



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
 TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

01/06/2014 16:27:03 GMT
 Page 1 of 1

Customer Name: Bel-Art Products-NJ Processing Facility: Salem Work Order # 1221325
 P.O.# MD 27442 Sales Order # 1069939
 23.5-50.0 kGy BelArt, Gamma Treatment Received Date/Time: 12/31/2013 14:57:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	118	CA	800371098	RESERVOIR W/LID, 12 CHANL, 5ML	J230230	0
102.000	164	CA	369060000(20)	STERILEWARE,SCOOP,PS,8 OZ	J229417	0
103.000	6	CA	369022006	STERILEWARE,SCOOP W-ANGLED HAN	J229489	0
104.000	5	CA	369042006	STERILEWARE,SCOOP W-ANGLED HAN	J229491	0
105.000	2	CA	369330002	STERILEWARE ECON SCOOP PS,83ML	J229801	0
106.000	175	CA	369440000(30)	STERILEWARE,SAMPLING SPOON,1 T	J229131	0
108.000	25	CA	369220000(30)	STERILEWARE,SCOOP SAMPLING SYS	J228703	0
109.000	3	CA	369330002	STERILEWARE ECON SCOOP PS,83ML	J229802	0
120.000	20	CA	800140113(20)	STERILIZED 1100cc PP SCOOP 24/	J229990	0
	518	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
500.00	Minimum Dose	23.5 kGy	50.0 kGy	24.4 KGY	Pass	RMACAULAY	01/06/2014 12:37:49 GMT
		Reason Code Test				RYAN MACAULAY	
500.00	Maximum Dose	23.5 kGy	50.0 kGy	37.3 KGY	Pass	RMACAULAY	01/06/2014 12:38:02 GMT
		Reason Code Test				RYAN MACAULAY	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EVA BUNDZA
 Reason: Work Order Completions

Date: 01/06/2014 16:26:26 GMT